



## CERTIFICATE OF ANALYSIS

Product:	GoBio™ Mini Cu-IDA 5 mL	Article number(s):	45 655 025 45 655 026 45 655 027 45 655 028
Lot number:	157529		
Storage solution:			20% ethanol

	Measured Value	Specification
Metal binding capacity before charging with Cu <sup>2+*</sup> :	64 µmol Cu <sup>2+</sup> /mL	50-70 µmol Cu <sup>2+</sup> /mL
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

\*Property measured on resin.

Release date (year-month-day): 2024-06-19

Expiry date (year-month): 2028-05

Manufacturing date (year-month): 2024-05

This product is manufactured in compliance with our ISO 9001:2015 certified quality system. Country of origin is Sweden.

Issued (year-month-day): 2024-06-19

By Markus Norén

DocuSigned by:

*Markus Norén*

6853712A40184AF...

Quality Assurance