



CERTIFICATE OF ANALYSIS

Product:	GoBio™ Mini affimAb 1 mL	Article number(s):	45 800 101 45 800 102 45 800 103 45 800 104
Lot number:	157553		
Storage solution:			20% ethanol

	Measured Value	Specification
Dynamic binding capacity for IgG*:	40.2 mg/mL	>40.0 mg/mL
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

*Property measured on resin.

Release date (year-month-day): 2024-06-18

Expiry date (year-month): 2028-05

Manufacturing date (year-month): 2024-05

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2024-06-18

By Markus Norén

DocuSigned by:

Markus Norén

6853712A40184AF...

Quality Assurance