
CERTIFICATE OF ANALYSIS

Product:	GoBio™ Mini S 1 mL	Article number(s):	45 200 101 45 200 102 45 200 103 45 200 104
Lot number:	157693		
Storage solution:	20% ethanol, 0.2 M sodium acetate		

	Measured Value	Specification
Ionic capacity*:	207 µmol H ⁺ /mL	180-250 µmol H ⁺ /mL
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

*Property measured on resin.

Release date (year-month-day): 2024-10-07

Expiry date (year-month): 2029-02

Manufacturing date (year-month): 2024-07

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2024-10-07

By Åsa Nilsson

DocuSigned by:
Åsa Nilsson
FD9074A2ACBC431...

Quality Assurance
