



CERTIFICATE OF ANALYSIS

Product:	GoBio™ Mini Co-NTA 5 mL	Article number(s):	45 655 135 45 655 136 45 655 137 45 655 138
Lot number:	157731		
Storage solution:			20% ethanol

	Measured Value	Specification
Metal binding capacity before charging with Co ²⁺ *	25 µmol Cu ²⁺ /mL	20-30 µmol Cu ²⁺ /mL
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

*Property measured on resin.

Release date (year-month-day): 2024-09-13

Expiry date (year-month): 2027-10

Manufacturing date (year-month): 2024-08

This product is manufactured in compliance with our ISO 9001:2015 certified quality system. Country of origin is Sweden.

Issued (year-month-day): 2024-09-13

By Markus Norén

DocuSigned by:

Markus Norén

8853712A40184AF...

Quality Assurance