



CERTIFICATE OF ANALYSIS

Product:	Article number(s):
WorkBeads™ 40/10 000 SEC	40 350 001
	40 350 003
Lot number: 157736	40 350 010
	40 350 050
Storage solution:	20% ethanol

	Measured Value	Specification
K _D (Ovalbumin):	0.71	0.65-0.85
Average particle size, Dv50:	48 µm	32-60 µm
Flow stability: (linear flow rate in a 10 x 300 mm column)*	≥10 cm/min	≥10 cm/min

*Property measured on base matrix.

Release date (year-month-day): 2024-08-28

Expiry date (year-month): 2029-08

Manufacturing date (year-month): 2024-08

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2024-09-06

By Markus Norén

DocuSigned by:



6853712A40184AF...

Quality Assurance
