



## CERTIFICATE OF ANALYSIS

Product:	GoBio™ Mini Ni-NTA 5 mL	Article number(s):	45 655 105 45 655 106 45 655 107 45 655 108 45 655 109
Lot number:	157755		
Storage solution:			20% ethanol

	Measured Value	Specification
Metal binding capacity before charging with Ni <sup>2+*</sup> :	25 µmol Cu <sup>2+</sup> /mL	20-30 µmol Cu <sup>2+</sup> /mL
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

\*Property measured on resin.

Release date (year-month-day): 2024-09-12


Expiry date (year-month): 2027-03

Manufacturing date (year-month): 2024-08

This product is manufactured in compliance with our ISO 9001:2015 certified quality system. Country of origin is Sweden.

Issued (year-month-day): 2024-09-12

By Markus Norén

DocuSigned by:  
  
6853712A40184AF...  
Quality Assurance