



---

## CERTIFICATE OF ANALYSIS

---

Product: **WorkBeads™ 40 SEC Cube** Article number(s): 40 301 050

Lot number: 157757

Storage solution: 20% ethanol

---

	<b>Measured Value</b>	<b>Specification</b>
K <sub>D</sub> (Thyroglobulin):	0.21	0.20-0.30
Average particle size, Dv50:	55 µm	32-60 µm
Flow stability: (linear flow rate in a 10 x 300 mm column)*	≥10 cm/min	≥10 cm/min

\*Property measured on base matrix.

---

Release date (year-month-day): 2024-09-19

Expiry date (year-month): 2029-09

Manufacturing date (year-month): 2024-09

---

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.  
Country of origin is Sweden.

Issued (year-month-day): 2024-10-28

By Markus Norén

DocuSigned by:

*Markus Norén*

6853712A40184AF...

Quality Assurance

---