



## CERTIFICATE OF ANALYSIS

Product: **WorkBeads™ 200 SEC SG** Article number(s): 20 300 350

Lot number: 157822

Storage solution: 20% ethanol

	Measured Value	Specification
K <sub>D</sub> (Ovalbumin):	0.66	0.62-0.75
K <sub>D</sub> (Thyroglobolin):	0.32	0.25-0.45
Average particle size, Dv50:	170 µm	170-210 µm
Particles within range 130-250 µm*	89 % (vol)	≥85 % (vol)

\*Property measured on base matrix.

Release date (year-month-day): 2024-10-16

Expiry date (year-month): 2029-10

Manufacturing date (year-month): 2024-10

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.  
Country of origin is Sweden.

Issued (year-month-day): 2024-10-16

By Markus Norén

DocuSigned by:

*Markus Norén*

8853712A40184AF...

Quality Assurance