



---

## CERTIFICATE OF ANALYSIS

---

Product:	GoBio™ Mini IDA 1 mL	Article number(s):	45 655 011 45 655 012 45 655 013 45 655 014
Lot number:	157859		
Storage solution:			20% ethanol

---

	Measured Value	Specification
Metal binding capacity*:	65 µmol Cu <sup>2+</sup> /mL	50-70 µmol Cu <sup>2+</sup> /mL
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

\*Property measured on resin.

---

Release date (year-month-day): 2024-11-15

Expiry date (year-month): 2025-12

Manufacturing date (year-month): 2024-10

---

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.  
Country of origin is Sweden.

Issued (year-month-day): 2024-11-15

By Markus Norén

DocuSigned by:

*Markus Norén*

6853712A40184AF...

Quality Assurance

---