



CERTIFICATE OF ANALYSIS

Product:	GoBio™ Mini Ni-IDA 5 mL	Article number(s):	45 655 005 45 655 006 45 655 007 45 655 008
Lot number:	157865		
Storage solution:			20% ethanol

	Measured Value	Specification
Metal binding capacity before charging with Ni ^{2+*} :	60 µmol Cu ²⁺ /mL	50-70 µmol Cu ²⁺ /mL
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

*Property measured on resin.

Release date (year-month-day): 2024-11-28

Expiry date (year-month): 2026-03

Manufacturing date (year-month): 2024-11

This product is manufactured in compliance with our ISO 9001:2015 certified quality system. Country of origin is Sweden.

Issued (year-month-day): 2024-11-28

By Markus Norén

DocuSigned by:

Markus Norén

6853712A40184AF...

Quality Assurance