



---

## CERTIFICATE OF ANALYSIS

---

Product:	GoBio™ Mini ACT 1 mL	Article number(s):	45 400 001 45 400 002 45 400 003 45 400 004
Lot number:	157909		
Storage solution:			20% ethanol

---

	Measured Value	Specification
Concentration of reactive groups*:	216 µmol/mL	>200 µmol/mL
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

\*Property measured on resin.

---

Release date (year-month-day): 2025-01-10

Expiry date (year-month): 2027-10

Manufacturing date (year-month): 2024-11

---

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.  
Country of origin is Sweden.

Issued (year-month-day): 2025-01-30

By Markus Norén

DocuSigned by:

*Markus Norén*

6853712A40184AF...

Quality Assurance

---