



CERTIFICATE OF ANALYSIS

Product:	GoBio™ Mini Co-IDA 1 mL	Article number(s):	45 655 031 45 655 032 45 655 033 45 655 034
Lot number:	157920		
Storage solution:			20% ethanol

	Measured Value	Specification
Metal binding capacity before charging with Co ²⁺ *	65 µmol Cu ²⁺ /mL	50-70 µmol Cu ²⁺ /mL
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

*Property measured on resin.

Release date (year-month-day): 2024-12-13


Expiry date (year-month): 2025-12

Manufacturing date (year-month): 2024-11

This product is manufactured in compliance with our ISO 9001:2015 certified quality system. Country of origin is Sweden.

Issued (year-month-day): 2024-12-13

By Markus Norén

DocuSigned by:

6853712A40184AF...
Quality Assurance