



CERTIFICATE OF ANALYSIS

Product:	Article number(s):
WorkBeads™ 40 Q	40 100 001
	40 100 002
	40 100 010
Lot number: 157986	40 100 050
	40 100 060
Storage solution:	20% ethanol

	Measured Value	Specification
Ionic capacity:	185 µmol Cl ⁻ /ml	180-250 µmol Cl ⁻ /ml
Average particle size, Dv50:	54 µm	32-60 µm
Flow stability: (linear flow rate in a 10 x 300 mm column)*	≥10 cm/min	≥10 cm/min

*Property measured on base matrix.

Release date (year-month-day): 2025-01-17

Expiry date (year-month): 2030-01

Manufacturing date (year-month): 2025-01

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2025-01-17

By Markus Norén

DocuSigned by:

Markus Norén

6853712A40184AF...

Quality Assurance
