

---

## CERTIFICATE OF ANALYSIS

---

Product: **WorkBeads™ 200 SEC SG** Article number(s): 20 300 350

Lot number: 157989

Storage solution: 20% ethanol

---

	<b>Measured Value</b>	<b>Specification</b>
K <sub>D</sub> (Ovalbumin):	0.67	0.62-0.75
K <sub>D</sub> (Thyroglobulin):	0.36	0.25-0.45
Average particle size, Dv50:	174 µm	170-210 µm
Particles within range 130-250 µm*	86 % (vol)	≥85 % (vol)

\*Property measured on base matrix.

---

Release date (year-month-day): 2025-02-11

Expiry date (year-month): 2030-01

Manufacturing date (year-month): 2025-01

---

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.  
Country of origin is Sweden.

Issued (year-month-day): 2025-02-19

By Markus Norén

---

Quality Assurance

---