



CERTIFICATE OF ANALYSIS

Product: **WorkBeads™ 200 SEC SG** Article number(s): 20 300 350

Lot number: 158124

Storage solution: 20% ethanol

	Measured Value	Specification
K _D (Ovalbumin):	0.63	0.62-0.75
K _D (Thyroglobolin):	0.29	0.25-0.45
Average particle size, Dv50:	188 µm	170-210 µm
Particles within range 130-250 µm*	93 % (vol)	≥85 % (vol)

*Property measured on base matrix.

Release date (year-month-day): 2025-03-27


Expiry date (year-month): 2030-03

Manufacturing date (year-month): 2025-03

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2025-03-27

By Markus Norén

DocuSigned by:

6853712A40184AF...
Quality Assurance