



CERTIFICATE OF ANALYSIS

Product: GoBio™ Mini S 1 mL
Lot number: 158183
Storage solution: 20% ethanol, 0.2 M sodium acetate

Article number(s):
45 200 101
45 200 102
45 200 103
45 200 104

	Measured Value	Specification
Ionic capacity*:	250 µmol H ⁺ /mL	180-250 µmol H ⁺ /mL
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

*Property measured on resin.

Release date (year-month-day): 2025-04-30

Expiry date (year-month): 2028-04

Manufacturing date (year-month): 2025-04

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2025-04-30

By Markus Norén

DocuSigned by:

Markus Norén

6853712A40184AF...

Quality Assurance
