



CERTIFICATE OF ANALYSIS

Product: **WorkBeads™ 40 SEC Cube** Article number(s): 40 301 050

Lot number: 158222

Storage solution: 20% ethanol

| | Measured Value | Specification |
|--|-----------------------|----------------------|
| K _D (Thyroglobulin): | 0.22 | 0.20-0.30 |
| Average particle size, Dv50: | 55 µm | 32-60 µm |
| Flow stability: (linear flow rate in a 10 x 300 mm column)* | ≥10 cm/min | ≥10 cm/min |

*Property measured on base matrix.

Release date (year-month-day): 2025-06-11

Expiry date (year-month): 2030-05

Manufacturing date (year-month): 2025-05

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2025-06-11

By Markus Norén

DocuSigned by:

Markus Norén

6853712A40184AF...

Quality Assurance
