



CERTIFICATE OF ANALYSIS

Product: **GoBio™ Mini IEX Screening Kit** Article number(s): **45 900 001**
Lot number: 158294

Kit consists of: Lot number:
GoBio Mini S 1 mL: 158183
GoBio Mini Q 1 mL: 157340
GoBio Mini DEAE 1 mL: 158033
GoBio Mini TREN 1 mL: 157950

Storage solution: 20% ethanol (0.2 M sodium acetate added to GoBio Mini S)

	Measured Value	Specification
Ionic capacity (GoBio Mini S)*:	250 µmol H ⁺ /mL	180-250 µmol H ⁺ /mL
Ionic capacity (GoBio Mini Q)*:	198 µmol Cl ⁻ /mL	180-250 µmol Cl ⁻ /mL
Ionic capacity (GoBio Mini DEAE)*:	127 µmol Cl ⁻ /mL	110-160 µmol Cl ⁻ /mL
Ionic capacity (GoBio Mini TREN)*:	165 µmol Cl ⁻ /mL	130-200 µmol Cl ⁻ /mL
Gap control:	Passed	No gaps
Maximum flow test:	Passed	Stable pressure
Visual control:	Passed	No visible defects

*Property measured on resin.

Release date (year-month-day): 2025-05-23

Expiry date (year-month): 2027-04

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2025-05-23

By Markus Norén

DocuSigned by:

Markus Norén

6853712A40184AF...

Quality Assurance