
CERTIFICATE OF ANALYSIS

Product:	WorkBeads™ Macro SEC	Article number(s):	40 370 001 40 370 003 40 370 010 40 370 050 40 370 060
Lot number:	158357		
Storage solution:			20% ethanol

	Measured Value	Specification
K _D (Thyroglobulin):	0.44	0.41-0.51
Average particle size, Dv50:	44 µm	32-60 µm
Flow stability: (linear flow rate in a 10 x 300 mm column)*	≥10 cm/min	≥10 cm/min

*Property measured on base matrix.

Release date (year-month-day): 2025-09-10

Expiry date (year-month): 2030-06

Manufacturing date (year-month): 2025-06

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2025-09-10

By Markus Norén

Quality Assurance
