



CERTIFICATE OF ANALYSIS

| | |
|--------------------|--------------------|
| Product: | Article number(s): |
| WorkBeads™ affimAb | 40 800 001 |
| | 40 800 002 |
| Lot number:158362 | 40 800 010 |
| | 40 800 050 |
| | 40 800 060 |
| Storage solution: | 20% ethanol |

| | Measured Value | Specification |
|--|----------------|---------------|
| Dynamic binding capacity for IgG: | 50.5 mg/ml | >40.0 mg/ml |
| Average particle size, Dv50: | 46 µm | 32-60 µm |
| Flow stability: (linear flow rate in a 10 x 300 mm column)* | ≥10 cm/min | ≥10 cm/min |

*Property measured on base matrix.

Release date (year-month-day): 2025-06-24

Expiry date (year-month):2028-06

Manufacturing date (year-month): 2025-06

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2025-06-24

By Markus Norén

DocuSigned by:

Markus Norén

6853712A40184AF...

Quality Assurance