



CERTIFICATE OF ANALYSIS

Product: GoBio™ Mini IDA His-Tag Screening Kit 1 mL
Article number(s): 45 700 001

Lot number: 158402

Kit consists of: Lot number:
GoBio Mini Ni-IDA 1 mL: 157804
GoBio Mini Cu-IDA 1 mL: 157911
GoBio Mini Co-IDA 1 mL: 157920
GoBio Mini Zn-IDA 1 mL: 157894

Storage solution: 20% ethanol

	Measured Value	Specification
Metal binding capacity before binding with metal ion:		
GoBio Mini Ni-IDA*:	60 µmol/mL	50-70 µmol Cu ²⁺ /mL
GoBio Mini Cu-IDA*:	65 µmol/mL	50-70 µmol Cu ²⁺ /mL
GoBio Mini Co-IDA*:	65 µmol/mL	50-70 µmol Cu ²⁺ /mL
GoBio Mini Zn-IDA*:	64 µmol/mL	50-70 µmol Cu ²⁺ /mL
Gap control:	Passed	No gaps
Maximum flow test:	Passed	Stable pressure
Visual control:	Passed	No visible defects

*Property measured on resin.

Release date (year-month-day): 2025-06-26

Expiry date (year-month): 2025-12

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2025-06-26

By Johanna Tawe

DocuSigned by:

78ADB347C12D40E...

Quality Assurance