



## CERTIFICATE OF ANALYSIS

Product:	Article number(s):
WorkBeads™ 40 DEAE	40 150 001
	40 150 002
	40 150 010
Lot number: 158405	40 150 050
	40 150 060
Storage solution:	20% ethanol

	Measured Value	Specification
Ionic capacity:	121 µmol Cl <sup>-</sup> /ml	110-160 µmol Cl <sup>-</sup> /ml
Average particle size, Dv50:	55 µm	32-60 µm
Flow stability: (linear flow rate in a 10 x 300 mm column)*	≥10 cm/min	≥10 cm/min

\*Property measured on base matrix.

Release date (year-month-day): 2025-07-01

Expiry date (year-month): 2030-06

Manufacturing date (year-month): 2025-06

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.  
Country of origin is Sweden.

Issued (year-month-day): 2025-07-01

By Markus Norén

DocuSigned by:

Markus Norén

6853712A40184AF...

Quality Assurance