



---

## CERTIFICATE OF ANALYSIS

---

Product:	WorkBeads™ Macro SEC	Article number(s):	40 370 001 40 370 003 40 370 010 40 370 050 40 370 060
Lot number:	158633		
Storage solution:			20% ethanol

---

	Measured Value	Specification
K <sub>D</sub> (Thyroglobulin):	0.42	0.41-0.51
Average particle size, Dv50:	43 µm	32-60 µm
Flow stability: (linear flow rate in a 10 x 300 mm column)*	≥10 cm/min	≥10 cm/min

\*Property measured on base matrix.

---

Release date (year-month-day): 2025-10-24

Expiry date (year-month): 2030-09

Manufacturing date (year-month): 2025-09

---

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.  
Country of origin is Sweden.

Issued (year-month-day): 2025-10-24

By Markus Norén

DocuSigned by:

*Markus Norén*

6853712A40184AF...

Quality Assurance

---