



CERTIFICATE OF ANALYSIS

Product:	GoBio™ Mini affimAb Edge 1 mL	Article number(s):	45 810 101 45 810 103 45 810 104
Lot number:	158641		
Storage solution:			20% ethanol

	Measured Value	Specification
Dynamic binding capacity for HIgG*:	75 mg/mL	≥70 mg/mL
Gap control:	Passed	No gaps
Maximum flow test:	Passed	Stable pressure
Visual control:	Passed	No visible defects

*Property measured on resin.

Release date (year-month-day): 2025-10-21

Expiry date (year-month): 2027-07

Manufacturing date (year-month): 2025-09

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is South Korea.

Issued (year-month-day): 2025-11-11

By Johanna Tawe

DocuSigned by:

78ADB347C12D40E...

Quality Assurance
