

---

## CERTIFICATE OF ANALYSIS

---

|                   |                    |                    |  |
|-------------------|--------------------|--------------------|--|
| Product:          | WorkBeads™ affimAb | Article number(s): | 40 800 001<br>40 800 002<br>40 800 010<br>40 800 050<br>40 800 060 |
| Lot number:       | 158701             |                    |  |
| Storage solution: |                    |                    | 20% ethanol  |

---

|  | Measured Value | Specification |
|--|----------------|---------------|
| Dynamic binding capacity for IgG:                              | 41.6 mg/ml     | >40.0 mg/ml   |
| Average particle size, Dv50:                                   | 46 µm          | 32-60 µm      |
| Flow stability:<br>(linear flow rate in a 10 x 300 mm column)* | ≥10 cm/min     | ≥10 cm/min    |

\*Property measured on base matrix.

---

Release date (year-month-day): 2025-10-27

Expiry date (year-month): 2028-10

Manufacturing date (year-month): 2025-10

---

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.  
Country of origin is Sweden.

Issued (year-month-day): 2025-10-27

By Markus Norén

---

Quality Assurance

---