



CERTIFICATE OF ANALYSIS

Product:	GoBio™ Mini NiMAC 5 mL	Article number(s):	45 655 315 45 655 316 45 655 317 45 655 318
Lot number:	158845		
Storage solution:			20% ethanol

	Measured Value	Specification
Metal binding capacity before charging with Ni ^{2+*} :	>60 µmol Cu ²⁺ /mL	>60 µmol Cu ²⁺ /mL
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

*Property measured on resin.

Release date (year-month-day): 2025-12-23

Expiry date (year-month): 2029-11

Manufacturing date (year-month): 2025-11

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2025-12-23

By Markus Norén

DocuSigned by:

Markus Norén

6853712A40184AF...

Quality Assurance
