



CERTIFICATE OF ANALYSIS

| | | | |
|-------------------|--------------------|--------------------|--|
| Product: | WorkBeads™ affimAb | Article number(s): | 40 800 001 40 800 002 40 800 010 40 800 050 40 800 060 |
| Lot number: | 158863 | | |
| Storage solution: | | | 20% ethanol |

| | Measured Value | Specification |
|--|----------------|---------------|
| Dynamic binding capacity for IgG: | 41.3 mg/ml | >40.0 mg/ml |
| Average particle size, Dv50: | 45 µm | 32-60 µm |
| Flow stability: (linear flow rate in a 10 x 300 mm column)* | ≥10 cm/min | ≥10 cm/min |

*Property measured on base matrix.

Release date (year-month-day): 2025-12-09

Expiry date (year-month): 2028-12

Manufacturing date (year-month): 2025-12

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2025-12-09

By Markus Norén

DocuSigned by:

Markus Norén

6853712A40184AF...

Quality Assurance
