



---

## CERTIFICATE OF ANALYSIS

---

Product:	WorkBeads™ affimAb	Article number(s):	40 800 001 40 800 002 40 800 010 40 800 050 40 800 060
Lot number:	158890		
Storage solution:			20% ethanol

---

	Measured Value	Specification
Dynamic binding capacity for IgG:	43.0 mg/ml	>40.0 mg/ml
Average particle size, Dv50:	47 µm	32-60 µm
Flow stability: (linear flow rate in a 10 x 300 mm column)*	≥10 cm/min	≥10 cm/min

\*Property measured on base matrix.

---

Release date (year-month-day): 2025-12-19

Expiry date (year-month): 2028-12

Manufacturing date (year-month): 2025-12

---

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.  
Country of origin is Sweden.

Issued (year-month-day): 2025-12-19

By Markus Norén

DocuSigned by:

*Markus Norén*

6853712A40184AF...

Quality Assurance

---