



CERTIFICATE OF ANALYSIS

| | | | |
|-------------------|--------------------|--------------------|--|
| Product: | GoBio™ Mini Q 5 mL | Article number(s): | 45 100 105 45 100 106 45 100 107 45 100 108 |
| Lot number: | 158944 | | |
| Storage solution: | | 20% ethanol | |

| | Measured Value | Specification |
|-------------------|------------------------------|----------------------------------|
| Ionic capacity*: | 182 µmol Cl ⁻ /mL | 180-250 µmol Cl ⁻ /mL |
| Gap control: | Passed | No gaps |
| Maximum flow test | Passed | Stable pressure |
| Visual control | Passed | No visible defects |

*Property measured on resin.

Release date (year-month-day): 2025-12-18

Expiry date (year-month): 2030-10

Manufacturing date (year-month): 2025-12

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2025-12-18

By Markus Norén

DocuSigned by:

Markus Norén

6853712A40184AF...

Quality Assurance
