



---

## CERTIFICATE OF ANALYSIS

---

|                    |                    |
|--------------------|--------------------|
| Product:           | Article number(s): |
| WorkBeads™ 100 Q   | 10 100 001         |
|                    | 10 100 002         |
|                    | 10 100 005         |
| Lot number: 159018 | 10 100 010         |
|                    | 10 100 050         |
|                    | 10 100 060         |
| Storage solution:  | 20% ethanol        |

---

|  | Measured Value               | Specification                    |
|--|------------------------------|----------------------------------|
| Ionic capacity:  | 180 µmol Cl <sup>-</sup> /ml | 140-200 µmol Cl <sup>-</sup> /ml |
| Average particle size, Dv50:                                   | 91 µm                        | 90-110 µm                        |
| Flow stability:<br>(linear flow rate in a 25 x 200 mm column)* | ≥1000 cm/hr                  | ≥1000 cm/hr                      |

\*Property measured on base matrix.

---

Release date (year-month-day): 2026-02-06

Expiry date (year-month): 2031-01

Manufacturing date (year-month): 2026-01

---

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.  
Country of origin is Sweden.

Issued (year-month-day): 2026-02-06

By Johanna Tawe

DocuSigned by:

78ADB347C12D40E...

Quality Assurance

---