



CERTIFICATE OF ANALYSIS

Product:	Article number(s):
WorkBeads™ 100 Q	10 100 001
	10 100 002
	10 100 005
Lot number: 159073	10 100 010
	10 100 050
	10 100 060
Storage solution:	20% ethanol

	Measured Value	Specification
Ionic capacity:	165 µmol Cl ⁻ /ml	140-200 µmol Cl ⁻ /ml
Average particle size, Dv50:	91 µm	90-110 µm
Flow stability: (linear flow rate in a 25 x 200 mm column)*	≥1000 cm/hr	≥1000 cm/hr

*Property measured on base matrix.

Release date (year-month-day): 2026-02-26

Expiry date (year-month): 2031-01

Manufacturing date (year-month): 2026-01

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2026-02-26

By Markus Norén

DocuSigned by:

Markus Norén

6853712A40184AF...

Quality Assurance