



CERTIFICATE OF ANALYSIS

Product:	GoBio™ Mini NiMAC 1 mL	Article number(s):	45 655 311 45 655 312 45 655 313 45 655 314
Lot number:	159167		
Storage solution:			20% ethanol

	Measured Value	Specification
Metal binding capacity before charging with Ni ^{2+*} :	110 µmol Cu ²⁺ /mL	>60 µmol Cu ²⁺ /mL
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

*Property measured on resin.

Release date (year-month-day): 2026-03-09

Expiry date (year-month): 2027-12

Manufacturing date (year-month): 2026-03

This product is manufactured in compliance with our ISO 9001:2015 certified quality system. Country of origin is Sweden.

Issued (year-month-day): 2026-03-09

By Johanna Tawe

DocuSigned by:

78ADB347C12D40E...

Quality Assurance