



CERTIFICATE OF ANALYSIS

Product:	GoBio™ Mini NTA 1 mL	Article number(s):	45 655 111 45 655 112 45 655 113 45 655 114
Lot number:	159225		
Storage solution:			20% ethanol

	Measured Value	Specification
Metal binding capacity*:	25 µmol Cu ²⁺ /mL	20-30 µmol Cu ²⁺ /mL
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

*Property measured on resin.

Release date (year-month-day): 2026-03-26

Expiry date (year-month): 2027-01

Manufacturing date (year-month): 2026-03

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2026-03-27

By Markus Norén

DocuSigned by:

Markus Norén

6853712A40184AF...

Quality Assurance
