



CERTIFICATE OF ANALYSIS

Product:	GoBio™ Mini Q 5 mL	Article number(s):	45 100 105 45 100 106 45 100 107 45 100 108
Lot number:	159293		
Storage solution:			20% ethanol

	Measured Value	Specification
Ionic capacity*:	188 µmol Cl ⁻ /mL	180-250 µmol Cl ⁻ /mL
Gap control:	Passed	No gaps
Maximum flow test	Passed	Stable pressure
Visual control	Passed	No visible defects

*Property measured on resin.

Release date (year-month-day): 2026-05-05

Expiry date (year-month): 2031-02

Manufacturing date (year-month): 2026-04

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2026-05-05

By Markus Norén

DocuSigned by:

Markus Norén

6853712A40184AF...

Quality Assurance
