
CERTIFICATE OF ANALYSIS

Product: **WorkBeads 40 Ni-NTA**

Article number: **40 651 001, 40 651 003, 40 651 010
40 651 011, 40 651 051, 40 651 060,
40 651 210**

Lot number: **600064-3020-035**


Product release date: **2017 DEC 31**

Expire date: **2022-12**

	Specifications	Results
Color	Turquoise blue	conforms
Agarose content	7.4-7.8 %	conforms
Particle size (Dv50)	32-60 μm	50.6 μm
Flow properties (linear flow in a 10x300 mm column)	≥ 10 cm/min	conforms
Metal ion capacity (tested with Cu^{2+})	20-30 $\mu\text{mol/ml}$	24 $\mu\text{mol/ml}$

Issued **2019 JUNE 27**

by



Lars Haneskog
Director Quality control