

CERTIFICATE OF ANALYSIS

Product: **WorkBeads 40 Ni-NTA PClin**
Article number: **CU40651201, CU40651211**
Lot number: **600064-3026-027**
Product release date: **2019 MAY**
Expire date: **2022-05**

	Specifications	Results
Color	Turquoise blue	conforms
Agarose content	7.4-7.8 %	conforms
Particle size (Dv50)	32-60 µm	49 µm
Flow properties (linear flow in a 10x300 mm column)	≥10 cm/min	conforms
Metal ion capacity (tested with Cu ²⁺)	20-30 µmol/ml	23.5 µmol/ml

Issued **2020 FEB 25**

by



Helena Pahlm
Quality Assurance Manager