
CERTIFICATE OF ANALYSIS

Product: **WorkBeads 40 Ni-NTA PClin**

Article number: **CU40651201, CU40651211**

Lot number: **600064-3026-036**


Product release date: **2019 OCT 20**

Expire date: **2022-10**

	Specifications	Results
Color	Turquoise blue	conforms
Agarose content	7.4-7.8 %	conforms
Particle size (Dv50)	32-60 µm	52 µm
Flow properties (linear flow in a 10x300 mm column)	≥10 cm/min	conforms
Metal ion capacity (tested with Cu ²⁺)	20-30 µmol/ml	25 µmol/ml

Issued **2020 MAR 02**

by



Helena Pahlm
Quality Assurance Manager