

## CERTIFICATE OF ANALYSIS

---

Product: **WorkBeads 40 NTA**  
Article number: **40 602 001, 40 602 003, 40 602 010**  
Lot number: **600064-3029-026**  
Product release date: **2020 MAY 09**  
Expire date: **2025-05**


---

	<b>Specifications</b>	<b>Results</b>
Color	Light beige	conforms
Agarose content	7.4-7.8 %	conforms
Particle size (Dv50)	32-60 $\mu\text{m}$	49 $\mu\text{m}$
Flow properties (linear flow in a 10 x300 mm column)	$\geq 10$ cm/min	conforms
Metal ion capacity (tested with $\text{Cu}^{2+}$ )	20-30 $\mu\text{mol/ml}$	25 $\mu\text{mol/ml}$

---

Issued **2020 MAY 19**

by



Helena Pahlm  
Quality Assurance Manager