
CERTIFICATE OF ANALYSIS

Product:	Article number(s):
WorkBeads™ 40 Ni-NTA	CU40651201
	CU40651211
Lot number: 600064-3035-026Ni	
Storage solution:	0.5% ProClin™ 300

	Measured Value	Specification
Metal binding capacity before charging with Ni ²⁺ .*	25 µmol Cu ²⁺ /ml	20-30 µmol Cu ²⁺ /ml
Average particle size, Dv50:	50 µm	32-60 µm
Flow stability: (linear flow rate in a 10 x 300 mm column)**	>10 cm/min	≥10 cm/min

* From supplier's Certificate of Analysis

**Property measured on base matrix.

Release date (year-month-day): 2022-06-10

Expiry date (year-month): 2027-02


Manufacturing date (year-month): 2022-02

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2022-11-21

By Philip Eriksson

DocuSigned by:



158C686C7E5E4A0...

Quality Assurance
