



## CERTIFICATE OF ANALYSIS

Product:	Article number(s):
WorkBeads™ 40 NTA	40 602 001
	40 602 003
Lot number: 600064-3035-032	40 602 010
	40 602 050
	40 602 060
Storage solution:	20% ethanol

	Measured Value	Specification
Metal binding capacity before charging with Ni <sup>2+</sup> .*	24,5 µmol Cu <sup>2+</sup> /ml	20-30 µmol Cu <sup>2+</sup> /ml
Average particle size, Dv50:	51 µm	32-60 µm
Flow stability: (linear flow rate in a 10 x 300 mm column)**	>10 cm/min	≥10 cm/min

\* From supplier's Certificate of Analysis

\*\*Property measured on base matrix.

Release date (year-month-day): 2022-12-05

Expiry date (year-month): 2026-04

Manufacturing date (year-month): 2022-04

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.  
Country of origin is Sweden.

Issued (year-month-day): 2024-09-12

By Markus Norén

DocuSigned by:

Markus Norén

6853712A40184AF...

Quality Assurance