
CERTIFICATE OF ANALYSIS

Product: **WorkBeads™ 40 Ni-NTA** Article number(s):
CU40651201
CU40651211

Lot number: 600064-3038-025Ni

Storage solution: 0.5% ProClin™ 300

	Measured Value	Specification
Metal binding capacity before charging with Ni ²⁺ .*	25 µmol Cu ²⁺ /ml	20-30 µmol Cu ²⁺ /ml
Average particle size, Dv50:	50 µm	32-60 µm
Flow stability: (linear flow rate in a 10 x 300 mm column)**	≥10 cm/min	≥10 cm/min

* From supplier's Certificate of Analysis

**Property measured on base matrix.

Release date (year-month-day): 2023-11-08

Expiry date (year-month): 2027-03

Manufacturing date (year-month): 2023-03

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2024-05-07

By Åsa Nilsson

DocuSigned by:

Åsa Nilsson

FD9074A2ACBC431...

Quality Assurance
