



## CERTIFICATE OF ANALYSIS

|                             |                    |
|-----------------------------|--------------------|
| Product:                    | Article number(s): |
| WorkBeads™ 40 NTA           | 40 602 001         |
|                             | 40 602 003         |
|                             | 40 602 010         |
| Lot number: 600064-3038-028 | 40 602 050         |
|                             | 40 602 060         |
| Storage solution:           | 20% ethanol        |

|   | Measured Value               | Specification                   |
|---|------------------------------|---------------------------------|
| Metal binding capacity before charging with Ni <sup>2+</sup> .* | 26 µmol Cu <sup>2+</sup> /ml | 20-30 µmol Cu <sup>2+</sup> /ml |
| Average particle size, Dv50:                                    | 51 µm                        | 32-60 µm                        |
| Flow stability:<br>(linear flow rate in a 10 x 300 mm column)** | ≥10 cm/min                   | ≥10 cm/min                      |

\* From supplier's Certificate of Analysis

\*\*Property measured on base matrix.

Release date (year-month-day): 2024-04-17

Expiry date (year-month): 2027-11

Manufacturing date (year-month): 2023-11

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.  
Country of origin is Sweden.

Issued (year-month-day): 2024-04-17

By Markus Norén

DocuSigned by:

Markus Norén

6853712A40184AF...

Quality Assurance