



CERTIFICATE OF ANALYSIS

Product:	Article number(s):
WorkBeads™ 40 NTA	40 602 001
	40 602 003
	40 602 010
Lot number: 600064-3044-031	40 602 050
	40 602 060
Storage solution:	20% ethanol

	Measured Value	Specification
Metal binding capacity before charging with Ni ²⁺ .*	24 µmol Cu ²⁺ /ml	20-30 µmol Cu ²⁺ /ml
Average particle size, Dv50:	52 µm	32-60 µm
Flow stability: (linear flow rate in a 10 x 300 mm column)**	≥10 cm/min	≥10 cm/min

* From supplier's Certificate of Analysis

**Property measured on base matrix.

Release date (year-month-day): 2025-08-19

Expiry date (year-month): 2029-08

Manufacturing date (year-month): 2025-08

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2025-08-19

By Markus Norén

DocuSigned by:

Markus Norén

6853712A40184AF...

Quality Assurance
