

## CERTIFICATE OF ANALYSIS

---

Product: **WorkBeads™ 40 SEC**  
Article number: **40 300 000**  
Lot number: **AL040**  
Manufacturing date: **2014-12-10**

---

	Measured values	Specifications
Average particle size ( $d_{50}$ )	45.6 $\mu\text{m}$	32-60 $\mu\text{m}$
Kd (Ovalbumin)	0.65	0.62-0.75
Flow properties (linear flow in a 10x300 mm column)	$\geq 10$ cm/min	$\geq 10$ cm/min

---

Issued 2017 August 30<sup>th</sup>  
by



Allan Simpson  
Quality Assurance Manager