

## CERTIFICATE OF ANALYSIS

---

Product: **WorkBeads™ 40 Q**

Article number: **40 100 001, 40 100 002, 40 100 010, 40 100 050,  
40 100 060,40 100 210, 40 100 250**

Lot number: **BB007**

Manufacturing date: **2015 FEB 24**

---

	<b>Measured values</b>	<b>Specifications</b>
Ionic capacity	190 µmol Cl <sup>-</sup> /mL	180-250 µmol Cl <sup>-</sup> /mL
Average particle size (Dv50)	44.2 µm	32-60 µm
Flow properties (linear flow in a 10x300 mm column)	>10 cm/min	≥10 cm/min

---

Issued 2019 APR 12  
by



---

Mats Johnson  
Chief Executive officer