

CERTIFICATE OF ANALYSIS

Product: **WorkBeads™ 40S**


Article number: **40 200 001, 40 200 002, 40 200 005, 40 200 010,
40 200 050, 40 200 060, 40 200 210, 40 200 250**

Lot number: **BD009**

Manufacturing date: **2016 APR 18**

	Measured values	Specifications
Ion exchange capacity	200 µmol Na ⁺ /mL	180 – 250 µmol Na ⁺ /mL
Average particle size (Dv50)	46.5 µm	32 - 60 µm
Flow properties (linear flow in a 10x300 mm column)	>10 cm/min	≥10 cm/min

Issued 2019 APR 12
by



Mats Johnson
Chief Executive officer