

CERTIFICATE OF ANALYSIS

Product: **WorkBeads™ 40/10 000 ACT**

Article number: **40 450 001, 40 450 003, 40 450 010, 40 450 050,
40 450 210, 40 450 250**

Lot number: **BH 023**

Product release date: **2015 JUL 19**

Expire date: **2021-03**

	Measured values	Specifications
Average particle size (Dv50):	46.7 µm	32 - 60 µm
Conc. of reactive groups:	283 µmol/ml	> 200 µmol/ml
Flow properties (linear flow in a 10x300 mm column):	>10 cm/min	≥ 10 cm/min

Issued 2019 JUL 04

by



Lars Haneskog
Director Quality control