

CERTIFICATE OF ANALYSIS

Product: **OptioBio™ 40Q 10x100**

Article number: **55 410 011**

Column lot number: **xxxx**

	Measured values	Specifications
Colour of chromatography bed		White to off white
Efficiency (Test method W077)	N/m	> 6350 N/m
Asymmetry (Test method W077)		0.75 - 1.50
Gap control (Test method W077)		No gap
Pressure at max recommended flow (450 cm/h, 6 ml/min)	MPa	Not specified
Bed height	mm	100 - 102 mm
Visual control (Test method W078)		Pass test

Issued 2017 June xx
by

Allan Simpson
Quality Assurance Manager