
CERTIFICATE OF ANALYSIS

Product: GoBio™ Prep 16x100 40 DEAE Article number: 55 415 021

Lot number: D02092209

Storage solution: 20% ethanol

	Measured Value	Specification
Ionic capacity*:	127 µmol Cl ⁻ /ml	110-160 µmol Cl ⁻ /ml
Efficiency:	5107 N/m	≥4000 N/m
Asymmetry:	1.3	0.7-1.4

*Property measured on resin.

Resin lot number: 155630

Release date (year-month-day): 2022-09-21 Expiry date (year-month): 2027-04
Manufacturing date (year-month): 2022-09

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.
Country of origin is Sweden.

Issued (year-month-day): 2022-09-21

By Anna Boman

DocuSigned by:



3C2F4EE14561428...

Quality Assurance