

---

## CERTIFICATE OF ANALYSIS

---

Product: GoBio™ Prep 16x100 40 DEAE      Article number: 55 415 021

Lot number: D02092210

Storage solution: 20% ethanol

---

	Measured Value	Specification
Ionic capacity*:	127 µmol Cl <sup>-</sup> /ml	110-160 µmol Cl <sup>-</sup> /ml
Efficiency:	5451 N/m	≥4000 N/m
Asymmetry:	1.1	0.7-1.4

\*Property measured on resin.

Resin lot number: 155630

---

Release date (year-month-day): 2022-09-21

Expiry date (year-month): 2027-04

Manufacturing date (year-month): 2022-09

---

This product is manufactured in compliance with our ISO 9001:2015 certified quality system.  
Country of origin is Sweden.

Issued (year-month-day): 2022-09-21

By Anna Boman

DocuSigned by:  
  
3C2E4EE14561428...

---

Quality Assurance